


# Biotage® Extrahera™ Precision and Accuracy Performance

This document details an example of the pipetting precision and accuracy data for four specified volumes across 24 (4 x 6) positions.

## Experimental Design

- » Precision and accuracy were tested by separately measuring the volume of 24 transfers of 50 µL, 100 µL, 500 µL and 1000 µL.
- » Water was dispensed sequentially (4 channels simultaneously) starting with column 1 (positions A1-D1) and ending with column 6 (positions A6-D6).
- » Performance was assessed by measuring the weights of 24 collection tubes before and after the addition of purified water.
- » Volume data was generated for each position using temperature corrected density values.



Because of the low volumes and weights involved, correction factors for temperature and evaporation rate were applied:

### 1. Correction for Evaporation

- a. The first 4 tubes (position A1-D1) were re-weighed after all 24 tubes had been measured.
- b. Evaporation rate was calculated as the average weight loss measured in these tubes divided by the number of minutes taken for 24 weight measurements.
- c. This was then multiplied by the time taken between sample addition and weighing for each tube.
- d. The mass of water calculated to have evaporated was then added to the unadjusted weight

### 2. Temperature Correction

- a. Water temperature was measured inside the Extrahera cabinet, to reflect that being dispensed.
- b. An accurate density was calculated for water at this temperature.

**Table 1.** Summary Data

Specified Volume (µL)	Average Measured Volume (µL) n=24	Standard Deviation	Specified Precision (% RSD)	Measured Precision (% RSD)	Specified Accuracy (Error ± %)	Measured Accuracy (Error ± %)
50	50.6	0.20	1.0	0.39	2.0	1.14
100	99.4	0.43	1.0	0.43	2.0	- 0.61
500	497	0.92	1.0	0.18	1.5	- 0.63
1000	997	1.65	1.0	0.17	1.0	- 0.34

## Precision and Accuracy with 50 $\mu$ L Transfer Volume

**Table 2.** Weight of empty tubes

	1	2	3	4	5	6
A	4.5901	4.6720	4.6421	4.6205	4.6821	4.5972
B	4.6666	4.5714	4.6431	4.6435	4.6114	4.6597
C	4.6391	4.6016	4.6528	4.6554	4.6622	4.6250
D	4.6918	4.6506	4.6739	4.6536	4.6498	4.5853

**Table 3.** Weight of tubes plus 50  $\mu$ L deionized water

	1	2	3	4	5	6
A	4.6405	4.7225	4.6923	4.6704	4.7326	4.6477
B	4.7169	4.6218	4.6933	4.6936	4.6617	4.7097
C	4.6892	4.6521	4.7031	4.7057	4.7121	4.6755
D	4.7420	4.7011	4.7243	4.7038	4.7003	4.6356

**Table 4.** Weight of water (g) unadjusted

	1	2	3	4	5	6
A	0.0504	0.0505	0.0502	0.0499	0.0505	0.0505
B	0.0503	0.0504	0.0502	0.0501	0.0503	0.0500
C	0.0501	0.0505	0.0503	0.0503	0.0499	0.0505
D	0.0502	0.0505	0.0504	0.0502	0.0505	0.0503

**Table 5.** Weight of water (g) corrected for evaporation between dispensing and weighing (1.929 mg/h)

	1	2	3	4	5	6
A	0.0505	0.0506	0.0503	0.0501	0.0507	0.0507
B	0.0504	0.0505	0.0503	0.0503	0.0505	0.0502
C	0.0502	0.0506	0.0504	0.0505	0.0501	0.0507
D	0.0503	0.0506	0.0506	0.0504	0.0507	0.0505

**Table 6.** Water volume (mL) allowing for a density of 0.99754 g/mL at 23 °C

	1	2	3	4	5	6
A	0.0506	0.0507	0.0504	0.0502	0.0508	0.0508
B	0.0505	0.0506	0.0505	0.0504	0.0506	0.0504
C	0.0503	0.0507	0.0506	0.0506	0.0502	0.0509
D	0.0504	0.0507	0.0507	0.0505	0.0508	0.0507

## Precision and Accuracy with 100 $\mu$ L Transfer Volume

**Table 7.** Weight of empty tubes

	1	2	3	4	5	6
A	4.5909	4.6719	4.6422	4.6205	4.6823	4.5974
B	4.6667	4.5715	4.6366	4.6437	4.6115	4.6506
C	4.6392	4.6019	4.6528	4.6552	4.6622	4.6250
D	4.6818	4.6508	4.6739	4.6493	4.6498	4.5853

**Table 8.** Weight of tubes plus 100  $\mu$ L deionized water

	1	2	3	4	5	6
A	4.6890	4.7710	4.7408	4.7191	4.7810	4.6959
B	4.7657	4.6706	4.7348	4.7424	4.7100	4.7492
C	4.7383	4.7008	4.7521	4.7549	4.7616	4.7246
D	4.7808	4.7499	4.7731	4.7485	4.7491	4.6844

**Table 9.** Weight of water (g) unadjusted

	1	2	3	4	5	6
A	0.0981	0.0991	0.0986	0.0986	0.0987	0.0985
B	0.0990	0.0991	0.0982	0.0987	0.0985	0.0986
C	0.0991	0.0989	0.0993	0.0997	0.0994	0.0996
D	0.0990	0.0991	0.0992	0.0992	0.0993	0.0991

**Table 10.** Weight of water (g) corrected for evaporation between dispensing and weighing (2.357 mg/h)

	1	2	3	4	5	6
A	0.0982	0.0992	0.0988	0.0988	0.0989	0.0988
B	0.0991	0.0992	0.0984	0.0989	0.0987	0.0989
C	0.0992	0.0990	0.0995	0.0999	0.0996	0.0999
D	0.0991	0.0993	0.0994	0.0994	0.0996	0.0994

**Table 11.** Water volume (mL) allowing for a density of 0.997422 g/mL at 23.5 °C

	1	2	3	4	5	6
A	0.0984	0.0995	0.0990	0.0990	0.0992	0.0990
B	0.0993	0.0995	0.0986	0.0992	0.0990	0.0991
C	0.0995	0.0993	0.0997	0.1002	0.0999	0.1001
D	0.0994	0.0995	0.0996	0.0997	0.0998	0.0997

## Precision and Accuracy with 500 $\mu$ L Transfer Volume

**Table 12.** Weight of empty tubes

	1	2	3	4	5	6
A	4.5906	4.6720	4.6423	4.6205	4.6822	4.5977
B	4.6668	4.5717	4.6362	4.6436	4.6113	4.6511
C	4.6393	4.6018	4.6526	4.6555	4.6620	4.6250
D	4.6817	4.6508	4.6739	4.6496	4.6497	4.5853

**Table 13.** Weight of tubes plus 500  $\mu$ L deionized water

	1	2	3	4	5	6
A	5.0840	5.1665	5.1366	5.1153	5.1770	5.0922
B	5.1608	5.0661	5.1314	5.1391	5.1070	5.1466
C	5.1341	5.0971	5.1485	5.1515	5.1585	5.1215
D	5.1763	5.1461	5.1699	5.1451	5.1461	5.0816

**Table 14.** Weight of water (g) unadjusted

	1	2	3	4	5	6
A	0.4934	0.4945	0.4943	0.4948	0.4948	0.4945
B	0.4940	0.4944	0.4952	0.4955	0.4957	0.4955
C	0.4948	0.4953	0.4959	0.4960	0.4965	0.4965
D	0.4946	0.4953	0.4960	0.4955	0.4964	0.4963

**Table 15.** Weight of water (g) corrected for evaporation between dispensing and weighing (3.429 mg/h)

	1	2	3	4	5	6
A	0.4935	0.4947	0.4945	0.4951	0.4951	0.4949
B	0.4941	0.4946	0.4954	0.4958	0.4960	0.4959
C	0.4949	0.4955	0.4962	0.4963	0.4969	0.4969
D	0.4948	0.4955	0.4963	0.4958	0.4968	0.4967

**Table 16.** Water volume (mL) allowing for a density of 0.9973 g/mL at 24 °C

	1	2	3	4	5	6
A	0.4949	0.4960	0.4959	0.4964	0.4965	0.4962
B	0.4955	0.4959	0.4968	0.4971	0.4974	0.4972
C	0.4963	0.4968	0.4975	0.4977	0.4982	0.4983
D	0.4961	0.4969	0.4976	0.4972	0.4981	0.4981

## Precision and Accuracy with 1000 $\mu$ L Transfer Volume

**Table 17.** Weight of empty tubes

	1	2	3	4	5	6
A	4.5901	4.6718	4.6421	4.6209	4.6822	4.5975
B	4.6667	4.5715	4.6365	4.6439	4.6113	4.651
C	4.6393	4.6021	4.6528	4.6556	4.6623	4.6251
D	4.6817	4.6506	4.6738	4.6494	4.6498	4.5853

**Table 18.** Weight of tubes plus 1000  $\mu$ L deionized water

	1	2	3	4	5	6
A	5.5808	5.6631	5.6334	5.6127	5.6752	5.5904
B	5.6582	5.5636	5.6296	5.6375	5.6060	5.6452
C	5.6321	5.5949	5.6474	5.6506	5.6584	5.6212
D	5.6736	5.6426	5.6671	5.6433	5.6447	5.5805

**Table 19.** Weight of water (g) unadjusted

	1	2	3	4	5	6
A	0.9907	0.9913	0.9913	0.9918	0.9930	0.9929
B	0.9915	0.9921	0.9931	0.9936	0.9947	0.9942
C	0.9928	0.9928	0.9946	0.9950	0.9961	0.9961
D	0.9919	0.9920	0.9933	0.9939	0.9949	0.9952

**Table 20.** Weight of water (g) corrected for evaporation between dispensing and weighing (4.071 mg/h)

	1	2	3	4	5	6
A	0.9909	0.9915	0.9916	0.9921	0.9934	0.9934
B	0.9917	0.9923	0.9934	0.9940	0.9951	0.9947
C	0.9930	0.9931	0.9949	0.9954	0.9965	0.9966
D	0.9921	0.9923	0.9936	0.9943	0.9954	0.9957

**Table 21.** Water volume (mL) allowing for a density of 0.997 g/mL at 25.2 °C

	1	2	3	4	5	6
A	0.9939	0.9945	0.9946	0.9951	0.9964	0.9963
B	0.9947	0.9953	0.9964	0.9970	0.9981	0.9977
C	0.9960	0.9961	0.9979	0.9984	0.9995	0.9996
D	0.9951	0.9953	0.9966	0.9973	0.9984	0.9987

## Ordering Information

Description	Part Number	Quantity
Biotage Extrahera	414001	1
Configuration Kit 96 Positions	414007	1
Configuration Kit 24 Positions	414008	1
Service Agreement - Priority - Extrahera	SER-EX-SAP	
Service Agreement - Limited - Extrahera	SER-EX-SAL	
First Year Maintenance Package Extrahera	SER-EX-FYMP	
Installation Extrahera	SER-EX-IN	
IV/OV on site Extrahera	SER-EX-IVOV	
Biotage Disposable Tips 1000 µL Clear	414141	10 x Pk/96
Column Rack 24 x 6 mL (tabless)	413640SP	1
Column Rack 24 x 3 mL	414174SP	1
Column Rack 24 x 1 mL	414169SP	1
Column Rack 96 x 1 mL (tabless)	414253SP	1
Sample Rack 16 x 100 mm 24 Positions	414254SP	1
Sample Rack 13 x 100 mm 24 Positions	414255SP	1
Sample Rack 12 x 75 mm 24 Positions	414256SP	1
Collection Rack 18 mm 24 Positions	414257SP	1
Vacuum Pump ME1C, 100 to 230VAC 50/60Hz	356330	1
96-well Collection plate, 1 mL	121-5202	Pk/50
96-well Collection plate, 2 mL	121-5203	Pk/50
12 x 75 mm Test Tubes, Uncapped	C44651	Pk/1000
13 x 100 mm Test Tubes, Uncapped	C40707	Pk/1000
16 x 100 mm Test Tubes, Uncapped	C40708	Pk/1000
Solvent Reservoir 25 mL	414045SP	Pk/25
Flow Through Plate 24	414203SP	1
Flow Through Plate 96	414201SP	1



### EUROPE

Main Office: +46 18 565900  
Toll Free: +800 18 565710  
Fax: +46 18 591922  
Order Tel: +46 18 565710  
Order Fax: +46 18 565705  
order@biotage.com  
Support Tel: +46 18 56 59 11  
Support Fax: + 46 18 56 57 11  
eu-1-pointsupport@biotage.com

### NORTH & LATIN AMERICA

Main Office: +1 704 654 4900  
Toll Free: +1 800 446 4752  
Fax: +1 704 654 4917  
Order Tel: +1 704 654 4900  
Order Fax: +1 434 296 8217  
ordermailbox@biotage.com  
Support Tel: +1 800 446 4752  
Outside US: +1 704 654 4900  
us-1-pointsupport@biotage.com

### JAPAN

Tel: +81 3 5627 3123  
Fax: +81 3 5627 3121  
jp\_order@biotage.com  
jp-1-pointsupport@biotage.com

### CHINA

Tel: +86 21 2898 6655  
Fax: +86 21 2898 6153  
cn\_order@biotage.com  
cn-1-pointsupport@biotage.com

To locate a distributor,  
please visit our website at  
www.biotage.com

#### Part Number: PPS360

© 2014 Biotage. All rights reserved. No material may be reproduced or published without the written permission of Biotage. Information in this document is subject to change without notice and does not represent any commitment from Biotage. E&OE. A list of all trademarks owned by Biotage AB is available at [www.biotage.com/legal](http://www.biotage.com/legal). Other product and company names mentioned herein may be trademarks or registered trademarks and/or service marks of their respective owners, and are used only for explanation and to the owners' benefit, without intent to infringe.